




TÜVRheinland®

ZERTIFIKAT  
CERTIFICATE

No.: 968/EL 868.00/12

<b>Product tested</b>	Strong Motion Recorder	<b>Certificate holder</b>	SYSCOM Instruments SA Rue de l'industrie 21 1450 Sainte-Croix Switzerland
<b>Type designation</b>	MRSK2002 with Sensor MS2002+ and Recorder MR2002-SM16-K		
<b>Codes and standards forming the basis of testing</b>	IEC 61508 Parts 1-7:2010 IEC 60780:1998 (in extracts) IEC 60880:2006		
<b>Intended application</b>	The device complies with the requirements of the relevant standards (hardware safety integrity SIL 2 acc. to IEC 61508 and systematic capability SIL 3 acc. to IEC 61508, Category A acc. to IEC 60880). The assessment was executed in accordance with IEC 60780.  Therefore the device can be used in applications up to SIL 2 (HFT=0) resp. SIL 3 (HFT=1) acc. to IEC 61508 and up to Category A acc. to IEC 61226.		
<b>Specific requirements</b>	The instructions of the associated Installation and Operating Manual shall be considered.		
This certificate is valid until 2017-12-14.			



The test report-no.: 968/EL 868.00/12 dated 2012-12-14 is an integral part of this certificate.

The holder of a valid licence certificate for the product tested is authorized to affix the test mark shown opposite to products, which are identical with the product tested.

TÜV Rheinland Industrie Service GmbH  
Bereich Automation  
Funktionale Sicherheit  
Am Grauen Stein, 51105 Köln

Köln, 2012-12-14

Certification Body for FS-Products

Dr.-Ing. Thorsten Gantevoort