



TÜVRheinland®

ZERTIFIKAT
CERTIFICATE

No.: 968/K 204.00/13

Product tested	Strong Motion Recorder	Certificate holder	SYSCOM Instruments SA Rue de l'industrie 21 1450 Sainte-Croix Switzerland
Type designation	NCC2002 with Sensor MS2005+ and Recorder MR2002-SM24-K		
Codes and standards forming the basis of testing	IEC 62138:2004 IEC 60780:1998 (in extracts)		
Intended application	The device complies with the requirements of the relevant standards (Category C acc. to IEC 62138). The assessment was executed in accordance with IEC 60780. Therefore the device can be used in applications up to Category C acc. to IEC 61226.		
Specific requirements	The instructions of the associated Installation and Operating Manual shall be considered.		
This certificate is valid until 2018-05-26.			

The issue of this certificate is based upon an examination, whose results are documented in report-no.: 968/K 204.00/13 dated 2013-05-26.

This certificate is valid only for products which are identical with the product tested. It becomes invalid at any change of the codes and standards forming the basis of testing for the intended application.

TÜV Rheinland Industrie Service GmbH
Bereich Automation
Funktionale Sicherheit
Am Grauen Stein, 51105 Köln

Köln, 2013-05-26

Certification Body for FS-Products

Dr.-Ing. Thorsten Gantevoort